

VERIFICATION OF CALIBRATION REPORT

FEB 25 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130121
Name	John Pernell	ID#	6260 Date 2/24/10
Agency	Fairbanks Airport Police & Fire	Phone #	474-2555
Instrument Location	5195 Brumbaugh Ave. Fairbanks, AK 99709		
Alco S/N		Target Value	High Pressure
X124670		.082	1050 psi.
Alco Test Values		1 st Alco	2 nd Alco
		.079	.079
Signature		John Pernell	

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of March 2010.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

FEBRUARY 24, 2010

OPERATOR'S NAME:

PERNELL/JOHN

OPERATOR'S NUMBER: 6260

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: FBK6

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .002 ✓

ALCO S/N: X124678

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.64 in ✓
ALCO TARGET .001 12:15
BLANK TEST .000 12:17
INTERNAL STANDARD VERIFIED 12:17
ALCO TV 29.64 in .079 ✓ 12:17
BLANK TEST .000 12:18
SUBJECT SAMPLE .000 12:18
BLANK TEST .000 12:19
ALCO TV 29.62 in .079 12:19
BLANK TEST .000 12:20

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121 ✓

FEBRUARY 24, 2010 ✓

TIME 12:27

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/29/00

HEATERS

SAMPLE CHAMBER: 45c

BREATH TUBE: 44c

BAROMETER: 29.64 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""'&()*+,-./0123456789:;<=>?@ABCEFGH
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmnop
qrstuvwxy z{|}~